

# S<sup>semper</sup> classic

# The established classic

A powdered, latex glove whose highly developed dipping process guarantees constant wall-thickness. Additionally, the micro-rough, nonslip surface provides good grip when handling instruments. Ideal tactile propertiesImage: Skin careImage: Skin care

CONTACT US! sempermed@harpsglobal.com www.sempermed.com HARPS Investment Asia Pte. Ltd. 9 Straits View, #08-10A Marina One West Tower Singapore 018937, Singapore Tel.: +65 6274 4861 · Fax: +65 6274 <u>6977</u> HARPS Europe GmbH Wiedner Guertel 9-13, 1100 Vienna, Austria Tel.: +43 1 9346348

# S semper med classic



#### **Ideal tactile properties**

The specially developed dipping process ensures even wall thickness distribution throughout the glove, from the fingers to the cuff. This results in ideal tactile sensitivity and less hand fatigue even during longer operations.



#### Skin care

The glove is powdered with pure cornstarch which makes it easy to don. The multiple washing process (leaching) makes the natural latex glove skin friendly and very comfortable to wear.



#### **Increased safety**

Sempermed surgical gloves are classified as Medical Device (MD) as well as Personal Protective Equipment (PPE). This means that the glove is allowed to be used for handling chemicals\*.





more product information

#### **Excellent grip and comfort**

Sensitivity and a safe grip are essential for a successful surgery. This glove has a micro-rough nonslip surface to provide a good grip when handling instruments.

### **Technical data**

Туре	sterile surgical glove for single use, powdered	
Color	natural white	
Size / Overall length as per EN 455-2	5½, 6 and 6½ / median 260 mm   7, 7½ and 8 / median 270 mm   8½ and 9 / median 280 mm	
Wall thickness measured in single layer in the palm area	min. 0.14 mm	
Barrier performance	AQL 0.65	
Material	natural rubber latex (NR)	
Glove shape	fully anatomical with rolled rim	
Force at break as per EN 455-2	≥9N	
Shelf life in original packaging if stored according to product specificatior	5 years M	
Sterilization	radiation STERILE R	
Labeling & conformity to standards	EN 455-1/-2/-3/-4, EN ISO 21420, EN ISO 374-1 (Type B), EN ISO 374-2, EN 16523-1, EN ISO 374-4, ISO 374-5, EN 421, EN 556, ISO 2859-1, ISO 15223-1, ASTM F1671/F1671M	
Purpose	Single-use medical device class IIa according to MD Directive 93/42/EEC Single-use protective glove – PPE category III (protection against chemical substances for limited time)* according to PPE Regulation (EU) 2016/425	



### Packaging

• Left and right glove with turned up cuff in inner pouch, sealed in peel pack.

- In dispenser carton with sterilization indicator: 70 pairs
- In transport carton with sterilization indicator: 420 pairs

## **Article numbers**

size 5½	826054520
size 6	826054600
size 6½	826054620
size 7	826054700
size 7½	826054720
size 8	826054800
size 8½	826054820
size 9	826054900

#### DM1/2 Version 1 09/2024 EN

Please note the Sempermed Chemical Resistance list is available on https://www.sempermed.com/chemical-resistance-brands **IMPORTANT NOTE**: The latest product information is available at www.sempermed.com under the «Products» tab. Failure to observe this information, in particular with regard to (chemical) resistance, frequency of use and tolerability of the gloves, can result in personal injury and/or material damage. Please note, glove changes may be necessary approximately every hour during certain procedures, depending on the specifics of such procedures and based on the surgeon's assessment. **HARPS is not liable for incorrect use of the gloves norwill HARPS be liable for improper storage and handling**. In case of **damage or unsual signs of wear and tear**, gloves shall not be **used**. In case of **damage or unsual signs of wear and tear**, gloves shall not be **used**. In case of **damage or unsual signs of wear and tear**, gloves shall not be **used**. In case of **damage or unsual signs of wear and tear**, gloves shall not be **used**. In case of **damage or unsual signs of wear and tear**, gloves shall not be **used**. In case of **damage or unsual signs of wear and tear**, gloves shall not be **used**. In case of **damage or unsual signs of wear and tear**, gloves **s**. This disclaimer is subject to mistakes, printing errors and amendments. \*Copyright 2024 HARPS